


Care and Maintenance of the
imp

Single Lever Clamp (Cat.# 903)



Lubrication:

Prior to each sterilization we recommend placing at least 2 drops of mineral oil or standard hospital grade lubricant behind the cam (Back side of the Clamp) while moving the lever between Open and Close.